



HELLENIC REPUBLIC
MINISTRY OF HEALTH
4th HEALTH REGION OF MACEDONIA AND
THRACE
GENERAL HOSPITAL OF THESSALONIKI
AGIOS PAVLOS
(Ethnikis Antistasis 161, 55134, Thessaloniki)

THESSALONIKI 21/04/2016

Refer. Number 4898

TO WHOM IT MAY CONCERN

Department: Procurement
Tel: 2313 304461-4
Fax: 2313 304452
Contact: Tzima D..
E-mail: promagpavlos@outlook.com

**SUBJECT « ANNOUNCEMENT – INVITATION TO SUBMIT TECHNICAL SPECIFICATIONS
AND STANDARDS»**

The General Hospital of Thessaloniki "AGIOS PAVLOS", having regard to:

1. The no. Ref . 2891 / 24.7.2015 letter of GCS on " Mandatory reporting process specifications and standards for all public health bodies
2. The fact that it is intended to be conducted an open public competition for «A NURSE CALL SYSTEM» in accordance with the Healthcare Supplies, Medicines and Services 2013 (Abbreviation in Greek "Π.Π.Υ.Φ.Υ." Program) on behalf of the the General Hospital of Thessaloniki "AGIOS PAVLOS", based on the following:

e-mail of the assigner authority	promagpavlos@outlook.com
Publication date of the current announcement	21-04-2016
Submission period of technical specifications and standards:	20 days, that is from 22-04-2016 till 12-05-2016
Service description	A NURSE CALL SYSTEM
CPV code:	33195000-3
Budget	30.000,00€
Account Code Number (abbreviation in Greek "KAE")	9749

INVITES ALL THE INTERESTED PARTIES

To submit to the Procurement Department technical specifications and standards of «A NURSE CALL SYSTEM», within an exclusive 20-days deadline, starting on the next day of the date of the current announcement,

The submission of the technical specifications and standards must take place in Greek or/and in English language, simultaneously in printed and electronic storage form (cd, dvd, usb memory stick) based on the following specifications:

- A. The nurse call system should be able to cover:
- Forty one (41) patient wards
 - Thirty one (31) bathrooms
 - Eight (8) nurse stations (one in each clinic)

- Eleven (11) stretcher bearers via wireless call
- B. Existing infrastructure (to be used):
- Each bed is next to a pull-down emergency call switch
 - All switch contacts are parallel to a single pair of wires, which ends up at the telephone distributor of each floor
 - Bathrooms (WCs) do not have wiring
 - The neurology clinic (3 wards and 4 WCs) do not have wiring
- C. System requirements (based on the current wiring infrastructure)
- Each nurse call will be made at the ward level and it will be noted at the nurse station equipment as the ward from which the call is made
 - Cancellation of a call should be made either by the cancellation button in each ward or by the nurse station equipment
 - Each ward should have only one cancellation button
 - Each nurse station equipment should serve only the clinic wards
 - Bathrooms should be equipped with wireless call systems
 - Each nurse station equipment should be able for at least eleven (11) wireless calls to DECT phones or pagers, which will be used by stretcher bearers
 - A data logging system should be included for recording calls and providing statistics on the response times per call.

Deputy Director
of G.H.of Thessaloniki AGIOS PAVLOS

Seridis Petros
MSc in Healthcare Management